

ISBN 0-626-17913-0

SANS 1384:2006

Edition 2

SOUTH AFRICAN NATIONAL STANDARD

Plastics containers for parenteral solutions

Published by Standards South Africa
1 dr lategan road groenkloof ☒ private bag x191 pretoria 0001
tel: 012 428 7911 fax: 012 344 1568 international code + 27 12
www.stansa.co.za
© Standards South Africa

standards
SouthAfrica
(a division of SABS)

SANS 1384:2006
Edition 2

Table of changes

Change No.	Date	Scope

Foreword

This South African standard was approved by National Committee StanSA SC 5140.36C, *Intravenous solutions, injectables and infusion sets – Plastics containers for parenteral solutions*, in accordance with procedures of Standards South Africa, in compliance with annex 3 of the WTO/TBT agreement.

This edition cancels and replaces the first edition (SABS 1384:1984).

Annexes A and B are for information only.

Contents

	Page
Foreword	
1 Scope	3
2 Normative references.....	3
3 Definitions	3
4 Requirements.....	5
4.1 Materials	5
4.1.1 General	5
4.1.2 Sterilizability	5
4.2 Physical requirements	5
4.2.1 Resistance to temperature stability, pressure and leakage	5
4.2.2 Strength of container when dropped.....	5
4.2.3 Transparency	5
4.2.4 Water vapour permeability	5
4.2.5 Particulate contamination	6
4.2.6 Cover	6
4.2.7 Access port	6
4.2.8 Strength of connection of insertion point and freedom from leakage	6
4.2.9 Injection point	6
4.2.10 Hanging device	6
4.2.11 Identification	6
4.3 Chemical requirements	7
4.3.1 Raw container or sheeting	7
4.3.2 Aqueous extract of the raw container or sheeting	7
4.4 Biological requirements	7
4.4.1 Bacterial endotoxins	7
4.4.2 Impermeability to micro-organisms	8
5 Physical tests	8
5.1 Resistance to temperature stability, pressure and leakage	8

SANS 1384:2006

Edition 2

Contents *(concluded)*

	Page
5.2 Strength of container when dropped	9
5.3 Transparency	9
5.4 Water vapour permeability	9
5.5 Particulate contamination	10
5.6 Penetration of insertion point	10
5.7 Strength of connection of insertion point and freedom from leakage	10
5.8 Sealing of injection point	11
5.9 Hanging device	11
5.10 Identification	11
6 Chemical tests	11
6.1 Residue on the ignition of the raw container or sheeting.....	11
6.2 Chemical tests on the aqueous extract (test solution)	12
6.2.1 Preparation of the test solution	12
6.2.2 Opalescence and colour	12
6.2.3 Acidity or alkalinity	12
6.2.4 Absorbance	12
6.2.5 Reducing substances	13
6.2.6 Determination of ammonia	13
6.2.7 Residue on evaporation	13
6.2.8 Heavy metals	13
7 Biological tests	13
7.1 Bacterial endotoxins	13
7.2 Impermeability to micro-organisms	14
8 Packing and marking	14
8.1 Packing	14
8.2 Marking	14
8.2.1 Infusion containers	14
8.2.2 Bulk packages	14
Annex A (informative) Further biological tests	15
Annex B (informative) Quality verification of plastics containers for parenteral solutions	15
Bibliography	16

Plastics containers for parenteral solutions

1 Scope

This standard covers requirements for plastics containers such as film bags or blow-moulded plastic bottles, having a total nominal capacity in the range 25 mL to 5 000 mL, for direct administration of aqueous infusion (injection) solutions. It does not cover containers for blood or blood products.

2 Normative references

The following referenced documents are indispensable for the application of this document. All normative documents are subject to revision and, since any reference to a normative document is deemed to be a reference to the latest edition of that document, parties to agreements based on this document are encouraged to take steps to ensure the use of the most recent editions of the normative documents indicated below. Information on currently valid national and international standards can be obtained from Standards South Africa.

2.1 Standards

ISO 15747, *Plastics containers for intravenous injection*.

2.2 Other publications

European Pharmacopoeia (EP).

United States Pharmacopoeia (USP).

3 Definitions

For the purposes of this document, the following definitions apply.

3.1

acceptable

acceptable to the authority administering this standard, or to the parties concluding the purchase contract, as relevant

3.2

access port

area of the infusion container consisting of the insertion point and the injection point, if applicable [ISO 15747]